

Last Calibration Date for the Reference Load Cell: _____

Next Calibration Date for the Reference Load Cell: _____

A	B	C	D	E	F	G	H	
Ram Pressure Calibration Log Sheet			Previous Pressure Calibration Value	Pressure Reading from the AFGB1 Display	Pressure Reading from the Load Cell Display	Correction Factor	New Pressure Calibration Value	
Initials	Date	Time	(unitless)	(kPa)	(kPa)	(unitless)	(unitless)	Comments
Example 1	23-JUL-99	12:03 PM	503	600	608	0.9868	496	<i>previous pressure readings were high</i>
Example 2	08-NOV-99	9:03 AM	496	600	598	1.0033	498	<i>previous pressure readings were low</i>

